

KF-A F10(SC)
**PARTICLE
FILTERING
HALF MASK**

White/Green/Black

EN 149:2001+A1:2009 FFP2 NR




STAY PROTECTED!

Our mask goes through rigorous testing and strictly meets EN 149:2001+A1:2009 guidelines. We follow the highest quality standards to make sure you get the protection you need.

KINGFA represents 200 three-dimensional mask production lines, with a capacity of 10 million pcs. per day.



Serves Worldwide



1 United State

Canton, Michigan

2 Germany

Wiesbaden

3 India

Chennai

4 Malaysia

Kuala Lumpur

5 China

Konggang, Tianjin
Mianyang, Sichuan
Chengdu, Sichuan
Wuhan, Hubei
Kunshan, Jiangsu

Qinpu, Shanghai
Ningbo, Zhejiang
Qingyuan, Guangdong
Guangzhou, Guangdong
Zhuhai, Guangdong

Inhouse Production



Personalized Service and
24/7 Online Tracking System

Increasing Efficiency




Raw Material Supply Chain
Management & Control

Short Leadtime



Overseas Distribution Centers

Risk Control & Management



SGS Inspection and
Quality Control

EN Intended Use

This product, which is normally used in the general working environment, is designed to provide reliable respiratory protection and against certain airborne particles and dust, block body fluids and so on. Please read these instructions thoroughly before use.

FIN Käyttötarkoitus

Tuotetta käytetään yleensä tavallisessa työympäristössä ja se on suunniteltu antamaan luotettava hengityssuojaus tiettyjä ilmassa olevia hiukkasia ja pölyä vastaan, estämään kehon nesteitä ja niin edelleen. Lue nämä ohjeet huolellisesti ennen käyttöä.

REF

Shelf life 2 years
Haltbarkeit 2 Jahre
Kesto aika 2 vuotta
Hållbarhet 2 år

KF-A F10(SC)

Color: ○ White ● Green ● Black




DE Bestimmungsgemäße Verwendung


Dieses Produkt, das normalerweise in der allgemeinen Arbeitsumgebung verwendet wird, soll einen zuverlässigen Atemschutz bieten und vor bestimmten luftgetragenen Partikeln und Staub schützen, Körperflüssigkeiten blockieren und so weiter.

SE Avsedd användning

Denna produkt, som normalt används i den allmänna arbetsmiljön, är utformad för att ge tillförlitligt andningskydd och mot vissa luftburna partiklar och damm, blockera kroppsvätskor och så vidare. Läs dessa instruktioner noggrant innan användning.

FEATURE


 PFE
@0.1 microns (%)
≥ 94


 BR Inhalation
30l/min(mbar)
≤ 0.7


LATEX FREE

 BR Exhalation
160l/min(mbar)
≤ 3.0



 Total Inward Leakage (TIL) (%)
≤ 8

 CO₂ Clearance (%)
≤ 1

 BR Inhalation
95l/min(mbar)
≤ 2.4

 Flame Spread
Class I

QUALITY CONTROL SYSTEM CERTIFICATION

ISO 9001 Certification

Certificate

Standard **ISO 9001:2015**

Certificate Registr. No. **01 100 1430282**

Certificate Holder:

GuangDong Kingfa Science and Technology Co., Ltd.

Unified Social Credit Code: 91441802077867032A

Registration Address: No. 28, Delong Road, Qingcheng Dist.

Shijiao Town, Qingyuan City, 511545 Guangdong, P. R. China

Operation Address: same as above

Scope:

Design and Manufacturing of Modified Plastics;
Design and Manufacturing of Masks and Non-Powered Air-
Purifying Particle Respirator

Validity:

Proof has been furnished by means of an audit that the
requirements of ISO 9001:2015 are met.

The certificate is valid from 2020-07-19 until 2023-07-18.
It remains valid subject to satisfactory surveillance audits.
First certification 2014

This certificate information can be searched on CNCA official
website <http://www.cnca.gov.cn>

2020-06-08



TÜV Rheinland Cert GmbH
Am Grauen Stein · 51105 Köln

QUALITY CONTROL SYSTEM CERTIFICATION

ISO 13485 Certification



Certificate

The Certification Body of
TÜV Rheinland LGA Products GmbH

hereby certifies that the organization
**GuangDong Kingfa Science and
Technology Co., Ltd.**
No.28, Delong Road, Qingcheng Dist.
Qingyuan City
511545 Guangdong
P.R. China

has established and applies a quality management system for medical devices
for the following scope:

**Design and Development, Manufacture and Distribution of
Disposable Medical Face Masks (non-sterile)**

Proof has been furnished that the requirements specified in

EN ISO 13485:2016

are fulfilled. The quality management system is subject to yearly surveillance.

Effective Date: 2020-07-13
Certificate Registration No.: SX 60150441 0001
An audit was performed. Report No.: 17054679 002
This Certificate is valid until: 2023-07-12

Certification Body



Date 2020-07-13

Fuxiu Sheng

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
Tel.: +49 221 806-1371 Fax: +49 221 806-3935 e-mail: cert-validity@de.tuv.com http://www.tuv.com/safety

QUALITY CONTROL SYSTEM CERTIFICATION

ISO 17025 Certification



**China National Accreditation Service for Conformity Assessment
LABORATORY ACCREDITATION CERTIFICATE
(Registration No. CNAS L2647)**

Analytical Center of Kingfa Science & Technology Co., Ltd.

(Legal Entity: Kingfa Science & Technology Co., Ltd.)

No.33, Kefeng Road, Science City, Hi-Tech Industrial Development Zone,
Guangzhou, Guangdong, China

is accredited in accordance with ISO/IEC 17025: 2017 General Requirements for the Competence of Testing and Calibration Laboratories(CNAS-CL01 Accreditation Criteria for the Competence of Testing and Calibration Laboratories) for the competence to undertake the service described in the schedule attached to this certificate.

The scope of accreditation is detailed in the attached schedule bearing the same registration number as above. The schedule forms an integral part of this certificate.

Effective Date: 2019-09-06

Expiry Date: 2023-07-02

Signed on behalf of China National Accreditation Service for Conformity Assessment

A handwritten signature in black ink, appearing to be the name of an official from the accreditation service.

China National Accreditation Service for Conformity Assessment (CNAS) is authorized by Certification and Accreditation Administration of the People's Republic of China (CNCA) to operate the national accreditation schemes for conformity assessment. CNAS is a signatory of the International Laboratory Accreditation Cooperation Mutual Recognition Arrangement (ILAC MRA) and the Asia Pacific Accreditation Cooperation Mutual Recognition Arrangement (APAC MRA).

The validity of the certificate can be checked on CNAS website at <http://www.cnas.org.cn/english/findanaccreditedbody/index.shtml>.

CE Certification

SGS

Certificate FI21/968210

GUANGDONG KINGFA SCI.& TECH.CO., LTD

NO.28 Delong Avenue, Shijiao Town,
Qingcheng District, Qingyuan City,
Guangdong Province,
China.

It is certified that the manufacturer's technical file and the PPE product detailed on
page 2 have been assessed and found to be in accordance with

Regulation (EU) 2016/425 Module B, EU type-examination

This certificate is valid from 08 February 2021 until 08 February 2026
1. Certified since 08 February 2021

Authorised by

SGS FIMKO OY, Notified Body 0598
Takomotie 8, FI-00380, Helsinki, Finland
t +358 9 696 361 f +358 9 692 5474 www.sgs.com

FINAS
Finnish Accreditation Service
S003 (EN ISO/IEC 17065)

Page 1 of 2



This document is issued by the Company subject to its General Conditions of Certification Services accessible at www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitations of liability, indemnification and jurisdictional issues established therein. The authenticity of this document may be verified at <http://www.sgs.com/en/certified-clients-and-products/certified-client-directory>. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Certificate FI21/968210, continued

**GUANGDONG KINGFA SCI.&
TECH.CO., LTD**

Regulation (EU) 2016/425
Module B, EU type-examination

Issue 1

PPE Product

KINGFA (logo) Model KF-A F10(SC) particle filtering half mask, consisting of a five layer (Polypropylene/Polyethylene terephthalate) disposable face mask (available in White/Green/Black), with Polypropylene covered iron nose clip, Polyethylene terephthalate/ Polyurethane in **Green/Black & White** ear loops and Polypropylene hook.

It is certified that the manufacturer's technical file and the above mentioned PPE have been assessed and found to meet the applicable Essential Health and Safety Requirements in Annex II of Regulation (EU) 2016/425 Personal Protective Equipment

The following have been applied:

EN 149: 2001 +A1:2009 (Respiratory protective devices- filtering half masks to protect against particles) device classification: FFP2 NR.

This certificate is issued on strict condition that appropriate checks on manufactured PPE, as detailed in Article 19(c) of the regulation are implemented and maintained while the model is in production.

Certification is based on technical file reference:

KF-A F10(SC), Dated 26.01.2021.

SGS Reference Number UK/CRS 242998

This certificate remains the property of SGS Fimko Oy to whom it must be returned on request.

SGS Certification



Certificate CN20/42082

The management system of

Guangdong KINGFA SCI. &TECH. Co., Ltd.

No.28, Delong Avenue, Shijiao Town, Qingcheng District,
Qingyuan City, Guangdong Province, 511545, P.R. China

has been assessed and certified as meeting the requirements of

Regulation (EU) 2016/425

Module D

For the following activities

Manufacture of FFP1/FFP2 Protective Respirator
(Note: all products marked CE0598 must have a valid EU Type
Examination Certificates issued under Module B or a valid EC type
examination certificate issued under Article 10 of the PPE Directive
89/686/EEC.)

This certificate is valid from 10 June 2020 until 9 June 2023
and remains valid subject to satisfactory surveillance audits.

Re certification audit due before 26 May 2023

Issue 1. Certified since 10 June 2020

Authorised by

SGS FIMKO OY, Notified Body 0598

Takomitie 8, FI-00380 Helsinki, Finland
t +358 9 696 361 f +358 9 692 5474 www.sgs.com

Page 1 of 1



This document is issued by the Company subject to its General Conditions of Certification Services accessible at www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitations of liability, indemnification and jurisdictional issues established therein. The authenticity of this document may be verified at <http://www.sgs.com/certified-clients-and-products/certified-client-directory>. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

TEST REPORTS



中国认可
国际互认
检测
TESTING
CNAS L0599

Test Report **SL52105220022801TX** Date: January 18, 2021 Page 1 of 10
 GUANGDONG KINGFA SCI.&TECH. CO., LTD.
 NO.28 DELONG AVE., SHIJIAO TOWN, QINGCHENG DISTRICT, QING YUAN, GUANGDONG, CHINA

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A) Particle filtering half mask

Sample Color : (A) white

Composition : (A) Polypropylene, Polyethylene terephthalate

Style No. : KF-A F10(SC)

Manufacturer : GUANGDONG KINGFA SCI.&TECH. CO., LTD.

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Jan 04, 2021

Testing Period : Jan 05, 2021 - Jan 18, 2021

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, for further details, please refer to the following page(s).

Conclusion:

Sample No.	Recommendation Level
(A)	FFP2 NR

Signed for and on behalf of
 SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd. 3rd Building, No. 889, Yishen Road, Xuhui District, Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
 Testing Center No. 889, Yishen Road, Xuhui District, Shanghai, China 200233 中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com



Test Report SL52105220022801TX Date: January 18, 2021 Page 2 of 10

Test Result

Personal Protective Equipment - Respiratory Protective Devices- Filtering Half Masks to Protect against Particles- Requirements, Testing, Marking
EN 149:2001+A1:2009

Clause 7.4 Packaging
(EN 149:2001+A1:2009 Clause 8.2)

Test Requirement	Results	Comment
Particle filtering half masks shall be offered for sale packaged in such a way that they are protected against mechanical damage and contamination before use.	Comply	Pass

Clause 7.5 Material
(EN 149:2001+A1:2009, Clause 8.2 & 8.3.1 & 8.3.2)

Test Requirement	Results	Comment
Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used.	Comply	Pass
After undergoing the conditioning described in 8.3.1 none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps.	Comply	
When conditioned in accordance with 8.3.1 and 8.3.2 the particle filtering half mask shall not collapse.	Comply	
Any material from the filter media released by the air flow through the filter shall not constitute a hazard or nuisance for the wearer.	Comply	

Clause 7.6 Cleaning and Disinfecting
(EN 149:2001+A1:2009, Clause 8.4 & 8.5 & 8.11)

Test Requirement	Results	Comment
If the particle filtering half mask is designed to be re-usable, the materials used shall withstand the cleaning and disinfecting agents and procedures to be specified by the manufacturer. With reference to 7.9.2, after cleaning and disinfecting the re-usable particle filtering half mask shall satisfy the penetration requirement of the relevant class.	Not applicable (Not designed to be re-usable)	N.A.

Clause 7.7 Practical Performance
(EN 149:2001+A1:2009, Clause 8.4)

Test Requirement	Results	Comment
The particle filtering half mask shall undergo practical performance tests under realistic conditions. These general tests serve the purpose of checking the equipment for imperfections that cannot be determined by the tests described elsewhere in this standard.	No imperfections	Pass



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/terms-and-conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/terms-and-conditions/terms-e-document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com
SGS-CSL (Shanghai) Technical Services (Shanghai) Co., Ltd. 3rd Building No. 888 Yishan Road, Xuhai District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
Testing Center: 中国·上海·徐汇区宜山路888号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com



Test Report

SL52105220022801TX

Date: January 18, 2021

Page 3 of 10

Clause 7.8 Finish of Parts

(EN 149:2001+A1:2009, Clause 8.2)

Test Requirement	Results	Comment
Parts of the device likely to come into contact with the wearer shall have no sharp edges or burrs.	No sharp edges or burrs	Pass

Clause 7.9.1 Total Inward Leakage

(EN 149:2001+A1:2009, Clause 8.5)

Test Requirement	Results	Comment
The total inward leakage consists of three components: face seal leakage, exhalation valve leakage (if exhalation valve fitted) and filter penetration. For particle filtering half masks fitted in accordance with the manufacturer's information, at least 46 out of the 50 individual exercise results (i.e. 10 subjects x 5 exercises) for total inward leakage shall be not greater than: 25% for FFP1, 11% for FFP2, 5% for FFP3 and, in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than: 22% for FFP1, 8% for FFP2, 2% for FFP3	Detail refer to Appendix 1	Meet FFP1, Meet FFP2

Appendix 1: Summarization of Test Data

Inward Leakage Test Data

Subject	Sample No.	Condition	Walk(%)	Head Side/side(%)	Head up/down(%)	Talk(%)	Walk(%)	Mean(%)
Zhou	1	A.R.	3.52	4.15	4.77	4.14	4.37	4.19
Luo	2	A.R.	5.16	5.58	4.74	5.86	5.34	5.34
Lu	3	A.R.	4.48	4.09	3.96	4.80	4.26	4.32
Wang	4	A.R.	3.22	3.74	3.82	3.20	2.08	3.21
Bao	5	A.R.	5.36	5.05	5.38	5.12	5.54	5.29
Ding	6	T.C.	3.90	3.00	3.47	3.14	3.73	3.45
Li	7	T.C.	5.98	5.19	5.92	5.74	5.12	5.59
Chen	8	T.C.	3.42	3.25	4.52	3.92	3.88	3.80
Song	9	T.C.	4.37	4.34	4.29	4.31	5.47	4.56
Ye	10	T.C.	5.65	5.33	5.66	5.69	6.14	5.69

Facial Dimension

Subject	Face length	Face Width	Face Depth	Mouth Width
Chen	125	150	120	58
Lu	115	132	107	48
Zhou	115	135	106	52
Li	125	130	107	46
Luo	125	136	100	43
Zheng	128	140	112	55
Wang	120	147	103	48
Song	120	140	100	50
Bao	130	134	104	50



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

SGS-CHINA Technical Services (Shanghai) Co., Ltd. | 3rd Building No.889, Yishan Road, Xuhui District Shanghai, China 200233 | t (86-21) 61402666 | f (86-21) 64958763 | www.sgs.com.cn
 Testing Center | 中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 | t (86-21) 61402666 | f (86-21) 64958763 | e sgs.china@sgs.com



Test Report **SL52105220022801TX** Date: **January 18, 2021** Page **4** of **10**

Ding	134	150	110	52
Liu	120	135	117	50
Ye	126	137	105	52

Clause 7.9.2 Penetration of Filter Material
(EN 149:2001+A1:2009, Clause 8.11 & EN 13274-7:2019)

Test Requirement			Results	Comment
The penetration of the filter of the particle filtering half mask shall meet the requirements of the following table.			Detail refer to Appendix 2	Meet FFP1, Meet FFP2, Meet FFP3
Classification	Maximum penetration of test aerosol			
	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min		
	%	%		
	max.	max.		
FFP1	20	20		
FFP2	6	6		
FFP3	1	1		

Appendix 2: Summarization of Test Data

Penetration of filter material

Aerosol	Condition	Sample No.	Penetration (%)
Sodium chloride test	As received	1	0.487
		2	0.313
		3	0.376
	Simulated wearing treatment	4	0.405
		5	0.413
		6	0.372
		7	0.604
	Mechanical strength +Temperature conditioned	8	0.576
		9	0.548
		10	0.153
Paraffin oil test	As received	11	0.127
		12	0.135
	Simulated wearing treatment	13	0.129
		14	0.147
		15	0.136
	Mechanical strength +Temperature conditioned	16	0.522
		17	0.582
		18	0.567

Flow conditioning : Single filter: 95.0 L/min



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

3rd Building, No. 889, Yiehan Road, Xuhui District, Shanghai, China 200233 t: (86-21) 61402666 f: (86-21) 64968763 www.sgs.com.cn
 中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t: (86-21) 61402666 f: (86-21) 64968763 e: sgs.china@sgs.com



Test Report SL52105220022801TX Date: January 18, 2021 Page 5 of 10

Clause 7.10 Compatibility with Skin
(EN 149:2001+A1:2009, Clause 8.4 & 8.5)

Test Requirement	Results	Comment
Materials that may come into contact with the wearer's skin shall not be known to be likely to cause irritation or any other adverse effect to health.	No irritation or any other adverse effect to health	Pass

Clause 7.11 Flammability
(EN 149:2001+A1:2009, Clause 8.6)

Test Requirement	Results	Comment
The material used shall not present a danger for the wearer and shall not be of highly flammable nature When tested, the particle filtering half mask shall not burn or not to continue to burn for more than 5 s after removal from the flame.	Detail refer to Appendix 3	Pass

Appendix 3: Summarization of Test Data

Flammability

Condition	Sample No.	Result
As received	1	NIL
	2	NIL
Temperature conditioned	3	NIL
	4	NIL

Clause 7.12 Carbon Dioxide Content of The Inhalation Air
(EN 149:2001+A1:2009, Clause 8.7)

Test Requirement	Results	Comment
The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1,0 % (by volume)	Detail refer to Appendix 4	Pass

Appendix 4: Summarization of Test Data

Carbon Dioxide Content of The Inhalation Air

Condition	Sample No.	Result
As received	1	0.5734
	2	0.5721
	3	0.5735
		Mean value: 0.57



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Docscheck@sgs.com
SGS-CTI (Shanghai) Technical Services (Shanghai) Co., Ltd. 3rd Building No. 889 Yishan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
Testing Center No. 889 Yishan Road, Xuhui District Shanghai, China 200233 中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com



Test Report

SL52105220022801TX

Date: January 18, 2021

Page 6 of 10

Clause 7.13 Head Harness

(EN 149:2001+A1:2009, Clause 8.4 & 8.5)

Test Requirement	Results	Comment
The head harness shall be designed so that the particle filtering half mask can be donned and removed easily.	Comply	Pass
The head harness shall be adjustable or self-adjusting and shall be sufficiently robust to hold the particle filtering half mask firmly in position and be capable of maintaining total inward leakage requirements for the device.	Comply	

Clause 7.14 Field of Vision

(EN 149:2001+A1:2009, Clause 8.4)

Test Requirement	Results	Comment
The field of vision is acceptable if determined so in practical performance tests.	Comply	Pass

Clause 7.15 Exhalation Valve(s)

(EN 149:2001+A1:2009, Clause 8.2 & 8.9.1 & 8.3.4 & 8.8)

Test Requirement	Results	Comment
(a) A particle filtering half mask may have one or more exhalation valve(s), which shall function correctly in all orientations.	Not applicable due to No exhalation valve	N.A.
(b) If an exhalation valve is provided it shall be protected against or be resistant to dirt and mechanical damage and may be shrouded or may include any other device that may be necessary for the particle filtering half mask to comply with 7.9.	Not applicable due to No exhalation valve	
(c) Exhalation valve(s), if fitted, shall continue to operate correctly after a continuous exhalation flow of 300 l/min over a period of 30 s.	Not applicable due to No exhalation valve	
(d) When the exhalation valve housing is attached to the faceblank, it shall withstand axially a tensile force of 10N applied for 10 s.	Not applicable due to No exhalation valve	



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

SGS-CSI (Shanghai) Technical Services (Shanghai) Co., Ltd.
Testing Center

3rd Building, No.889, Yishan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com



Test Report

SL52105220022801TX

Date: January 18, 2021

Page 7 of 10

Clause 7.16 Breathing Resistance
(EN 149:2001+A1:2009, Clause 8.9)

Test Requirement				Results	Comment
The breathing resistances apply to valved and valveless particle filtering half masks and shall meet the requirements of the following table.				Detail refer to Appendix 5	Meet FFP1, Meet FFP2, Meet FFP3
Classification	Maximum permitted resistance (mbar)				
	Inhalation		Exhalation		
	30 l/min	95 l/min	160 l/min		
FFP1	0.6	2.1	3.0		
FFP2	0.7	2.4	3.0		
FFP3	1.0	3.0	3.0		

Appendix 5: Summarization of Test Data

Breathing resistance (mbar)

As received	Flow rate(l/min)	1					2					3				
		A	B	C	D	E	A	B	C	D	E	A	B	C	D	E
Inhalation	30	0.4	0.4	0.4	0.4	0.3	0.3	0.4	0.3	0.3	0.4	0.4	0.4	0.4	0.3	0.4
	95	1.5	1.5	1.5	1.5	1.4	1.4	1.4	1.5	1.5	1.5	1.4	1.5	1.4	1.5	1.5
Exhalation	160	2.2	2.2	2.1	2.2	2.2	2.2	2.2	2.1	2.1	2.2	2.1	2.2	2.2	2.1	2.2
Simulated wearing treatment	Flow rate(l/min)	4					5					6				
		A	B	C	D	E	A	B	C	D	E	A	B	C	D	E
Inhalation	30	0.4	0.3	0.4	0.4	0.3	0.3	0.4	0.4	0.4	0.4	0.3	0.3	0.4	0.4	0.4
	95	1.5	1.4	1.5	1.5	1.4	1.4	1.5	1.5	1.5	1.4	1.5	1.5	1.4	1.5	1.4
Exhalation	160	2.1	2.1	2.2	2.2	2.2	2.1	2.2	2.2	2.2	2.2	2.1	2.1	2.1	2.1	2.2
Temperature conditioned	Flow rate(l/min)	7					8					9				
		A	B	C	D	E	A	B	C	D	E	A	B	C	D	E
Inhalation	30	0.3	0.4	0.3	0.3	0.4	0.4	0.3	0.3	0.4	0.3	0.3	0.4	0.4	0.3	0.3
	95	1.4	1.4	1.5	1.5	1.4	1.4	1.4	1.5	1.5	1.4	1.4	1.4	1.4	1.4	1.5
Exhalation	160	2.1	2.2	2.1	2.1	2.2	2.2	2.1	2.1	2.1	2.1	2.1	2.2	2.2	2.1	2.1

A: facing directly ahead; B: facing vertically upwards; C: facing vertically downwards; D: lying on the left side; E: lying on the right side



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com
SGS-CTE (Shanghai) Technical Services Co., Ltd. | 13' Building, No. 889, Yishan Road, Xuhai District, Shanghai, China 200233 | t (86-21) 61402666 | f (86-21) 64858763 | www.sgs.com.cn
Testing Center for Particle Services | 中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 | t (86-21) 61402666 | f (86-21) 64858763 | e sgs.china@sgs.com



Test Report

SL52105220022801TX

Date: January 18, 2021

Page 8 of 10

Clause 7.17 Clogging
(EN 149:2001+A1:2009, Clause 8.9 & 8.10)

Test Requirement	Results	Comment																				
<p>Clause 7.17.2 Breathing resistance Valved particle filtering half masks: After clogging the inhalation resistances shall not exceed: FFP1: 4 mbar, FFP2: 5 mbar, FFP3: 7 mbar at 95L/min continuous flow The exhalation resistance shall not exceed 3 mbar at 160 L/min continuous flow.</p> <p>Valveless particle filtering half masks: After clogging the inhalation and exhalation resistances shall not exceed: FFP1: 3 mbar, FFP2: 4 mbar, FFP3: 5 mbar at 95L/min continuous flow</p>	Optional for single shift device only	N.A.																				
<p>Clause 7.17.3 Penetration of filter material All types (valved and valveless) of particle filtering half masks claimed to meet the clogging requirement shall also meet the requirements.</p> <table border="1"> <thead> <tr> <th rowspan="2">Classification n</th> <th colspan="2">Maximum penetration of test aerosol</th> </tr> <tr> <th>Sodium chloride test 95 l/min</th> <th>Paraffin oil test 95 l/min</th> </tr> </thead> <tbody> <tr> <td></td> <td>%</td> <td>%</td> </tr> <tr> <td></td> <td>max.</td> <td>max.</td> </tr> <tr> <td>FFP1</td> <td>20</td> <td>20</td> </tr> <tr> <td>FFP2</td> <td>6</td> <td>6</td> </tr> <tr> <td>FFP3</td> <td>1</td> <td>1</td> </tr> </tbody> </table>	Classification n	Maximum penetration of test aerosol		Sodium chloride test 95 l/min	Paraffin oil test 95 l/min		%	%		max.	max.	FFP1	20	20	FFP2	6	6	FFP3	1	1	Optional for single shift device only	N.A.
Classification n		Maximum penetration of test aerosol																				
	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min																				
	%	%																				
	max.	max.																				
FFP1	20	20																				
FFP2	6	6																				
FFP3	1	1																				

Clause 7.18 Demountable Parts
(EN 149:2001+A1:2009, Clause 8.2)

Test Requirement	Results	Comment
All demountable parts (if fitted) shall be readily connected and secured, where possible by hand	No demountable parts	N.A.

Test	Uncertainty
Total inward leakage	3.4%
Penetration of filter material	4.8%
Carbon dioxide content of the inhalation air	3.9%
Breathing resistance (30L/min)	5.9%
Breathing resistance (95L/min)	4.9%
Breathing resistance (160L/min)	4.3%



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

SGS-CSI (Shanghai) Technical Services (Shanghai) Co., Ltd. 3rd Building, No. 889, Yeshan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgsgroup.com.cn
 Testing Center: 中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com



Test Report

SL52105220022801TX

Date: January 18, 2021

Page 9 of 10

Sample Photo



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
SGS-CSTC Shanghai Technical Services Co., Ltd. 3rd Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
Testing Center for Physical Properties of Textiles 中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com


SGS

Test Report

SL52105220022801TX

Date: January 18, 2021

Page 10 of 10



End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

3rd Building, No. 889 Yishan Road, Xuhai District, Shanghai, China 2000233 t (86-21) 61402666 f (86-21) 64958763 www.sgsgroup.com.cn
 中国·上海·徐汇区宜山路889号3号楼 邮编: 2000233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com

EU DECLARATION OF CONFORMITY

EU DECLARATION OF CONFORMITY

We hereby declare that the following items complies with the EU regulation 2016/425/EEC Personal Protective Equipment (PPE).

This declaration of conformity is issued under the sole responsibility of the manufacturer:

[Manufacturer and address]

GUANGDONG KINGFA SCI.&TECH. CO., LTD.
No.28 Delong Avenue, Shijiao Town, Qingcheng District,
Qingyuan City, Guangdong Province, China.

[Product name] Particle Filtering half mask

[Model/ Serial No.] KF-A F10(SC) FFP2 NR

[Applicable Regulation/Standard(s):]

This PPE is in conformity with the the relevant Union Harmonisation Legislation Regulation (EU) 2016/425 and with the harmonized standards EN149:2001 + A1:2009

[Conformity Assessment Procedure]

EU type-examination (module B) certificate FI21/968210.
Issued By (Notified Body name/number):SGS Fimko Oy (NB 0598).
Notified body address : Takomotie 8, FI-00380, Helsinki, Finland
Certificate of conformity to type based on quality assurance of the production process (module D) is issued by SGS FIMKO OY (NB 0598)

[EC representative:]

Lotus NL B.V.
Koningin Julianaplein 10, 1e Verd, 2595AA, The Hague, Netherlands.

[Signed for and on behalf of]

GUANGDONG KINGFA SCI.&TECH. CO., LTD.

[Place of issue] Qingyuan City, Guangdong Province, China

[Date of signature] 2021-02-09

[Title of signatory] General Manager

[Signature]



EU-KONFORMITÄTSERKLÄRUNG

Wir erklären hiermit, dass die folgenden Punkte der EU-Verordnung 2016/425/EWG über persönliche Schutzausrüstung (PSA) entsprechen.

Diese Konformitätserklärung wird unter alleiniger Verantwortung des Herstellers erstellt.

[Name und Anschrift]

GUANGDONG KINGFA SCI.&TECH. CO., LTD.
No.28 Delong Avenue, Shijiao Town, Qingcheng District,
Qingyuan City, Guangdong Province, China.

[Produktidentifikation] Partikelfilter-halbmaske

[Modell. Nr.] KF-A F10(SC) FFP2 NR

[Anwendbare Verordnung/Norm(en)]

Diese PSA steht in Einklang mit der einschlägigen Harmonisierungsrechtsvorschriften Verordnung (EU) 2016/425 und, wenn dies der Fall ist, mit den harmonisierten Normen EN149:2001 + A1:2009.

[Konformitätsbewertungsverfahren]

EG-Baumusterprüfbescheinigung FI21/968210

Ausgestellt von (Name / Nummer der benannten Stelle): SGS Fimko Oy (NB 0598)

Angemeldete Vertretung: Takomotie 8, FI-00380, Helsinki, Finland

Diese PSA unterliegt der Konformität von Modul D mit dem Typ basierend auf der Qualität.

Die Gewährleistung des Produktionsprozesses erfolgt durch das SGS FIMKO OY (NB 0598).

[Vertreter der EC]

Lotus NL B.V.

Koningin Julianaplein 10, 1e Verd, 2595AA, The Hague, Netherlands.

[Firmenname] GUANGDONG KINGFA SCI.&TECH. CO., LTD.

[Ausstellungsort] Qingyuan City, Guangdong Province, China

[Datum der Ausstellung] 2021-02-09

[Position des Unterzeichners] Geschäftsführer

[Unterschrift]



PACKING INFORMATION



White/Green/Black

Inner Bag

- Size:125*140mm



1
pc/bag



Paper Box

- Size:140*35*125mm



6
pcs/box



Carton

- 200 boxes/carton
- Size:575*370*650mm
- Gross weight:15500±500g




1200
pcs/caton



The actual product you receive may be (slightly) different from this photo please make the object as the standard

STORAGE AND DISPOSAL



PACKING INFORMATION



White/Green/Black

Paper Box

- Size:140*40*121mm



10
pcs/box



Carton

- 120 boxes/carton
- Size:510*435*440mm
- Gross weight:13500±500g





1200
pcs/caton



The actual product you receive may be (slightly) different from this photo please make the object as the standard

STORAGE AND DISPOSAL



PACKING INFORMATION



White

Paper Box

- Size:135*120*121mm



30
pcs/box



Carton

- 40 boxes/carton
- Size:625*285*505mm
- Gross weight:12500±500g




1200
pcs/carton



The actual product you receive may be (slightly) different from this photo please make the object as the standard

STORAGE AND DISPOSAL



KINGFA | MEDICAL
P R O T E C T I N G P E O P L E

Contact your KINGFA representative for more information

**GUANGDONG KINGFA
SCI. & TECH. CO., LTD.**



NO.28, Delong Avenue, Shijiao Town, Qingcheng District,
Qingyuan City, Guangdong Province, China



+86 (020) 6681 8888



medical@kingfa.com



<http://www.kingfa.com.cn>